

LabSen 371 Flat pH Electrode User Manual

LabSen electrochemical sensors are premium pH electrodes with manufacturing technology and key components imported from Switzerland. LabSen 371 plastic flat pH electrode adopts a flat membrane and PTFE junction structure. It is suitable for wet solid and semi-solid surface, such as paper, skin, textile and leather.

This probe has following features:

- Anti-corrosion POM (Polyoxymethylene) shaft, have good mechanical performance.
- Blue gel inner solution, does not flow and will not cause bubble.
- Long life reference system, has better stability and service life.

1 Technical Data

Measuring Range	(0 ~ 14) pH	Electrolyte	Gel
Temperature Range	(0 ~ 80) °C	Soaking Solution	3M KCl
Shaft Material	POM	Electrode Dimension	(Ø12×105) mm
Membrane Type	Flat	Membrane Resistance	<500MΩ
Reference	Long Life	Connector	BNC
Junction	PTFE	Cable	Ø3mm × 1m

2 Usage and Maintenance

- 2.1 Testing surface needs to be clean and wet, such as wetting the surface of skin or textiles with distilled water.
- 2.2 Connect BNC plug to the input on the pH meter.
- 2.3 When measuring, please unscrew the bottle cap, pull out the electrode and rinse it with deionized/distilled water. After using, please put the electrode back into the bottle and screw tight the cap.
- 2.4 The electrode's measuring tip should be soaked in the soaking bottle containing certain amount of storage solution to keep the membrane hydrated and junction unblocked. Clean the bottle and replace the storage solution if the storage solution gets turbid and mildewed. The electrode should never be soaked in pure water or buffer solution for long.
- 2.5 The connector of the electrode should always be clean and dry. If it gets contaminated, please clean it with medical cotton and absolute alcohol and blow dry it to prevent the short circuit of the electrode and slow reaction of electrode.
- 2.6 Every pH electrode will eventually age and fail. The typical service life of Apera pH electrodes is 12-24 months depending on the frequency of usage and how well you keep it clean and properly stored. We recommend replacing your electrode every 12-18 months to ensure the best accuracy.

3 Warranty

We warrant this electrode free from defects in material and workmanship and agrees to repair or replace free of charge, at option of APERA INSTRUMENTS any malfunctioned or damaged product attributable to responsibility of APERA INSTRUMENTS for a period of **six months**. Warranty period is the time limit to provide free service for the products purchased by customers, not the service life of the tester or electrodes.

This limited warranty does not cover any damages due to:

transportation; storage; improper use; failure to follow the product instructions or to perform any preventive maintenance; modifications; combination or use with any products, materials, processes, systems or other matter not provided or authorized in writing by us; unauthorized repair; normal wear and tear; or external causes such as accidents, abuse, or other actions or events beyond our reasonable control.

APERA INSTRUMENTS (Europe) GmbH

Address: Wilhelm-Muthmann-Str.18

42329 Wuppertal, Germany

Phone: +49 202 51988998

Website: www.aperainst.de

Email: info@aperainst.de