Der APERA INSTRUMENTS EC60-DE Ersatz -Sensor (Elektrode) für die Series 60-Z ist kompatibel mit:

● **EC60-Z** ●**PC60-Z**

- Nehmen Sie die Sensorkappe ab.
- Schrauben Sie den Ring am Sensor ab und lösen Sie den Sensor.
- Stecken Sie den neuen Sensor an das Gerät. Beachten Sie dabei die korrekte Ausrichtung.
- Schrauben Sie den Ring wieder an.



**Video-Tutorials finden Sie unter www.aperainst.de oder auf YouTube (Apera Instruments Europe Official Channel).

APERA INSTRUMENTS (Europe) GmbH gewährt auf dieses Produkt eine Garantie von Sechs Monate. Die Herstellergarantie beginnt ab dem Erstkaufdatum durch den ersten Endkunden (Rechnungsdatum). Die Garantie umfasst die fehlerfreie Funktion des Produktes. Sollten während der Garantiezeit Mängel des Produktes herausstellen, die auf Herstellung- oder Verarbeitungsfehlern beruhen, so wird APERA INSTRUMENTS (Europe) GmbH das Produkt oder den defekten Teil kostenfrei reparieren oder (nach Ermessen) ersetzen. Ausgenommen von der Garantie sind insbesondere Schäden, die durch unsachgemäßen Gebrauch, Missachtung der Bedienungsanleitung, nicht autorisierte Reparaturen und Modifikationen sowie Verschleiß entstanden sind.

APERA INSTRUMENTS (Europe) GmbH
Wilhelm-Muthmann-Straße 15, 42329 Wuppertal, Germany
Tel. +49 202 51988998
info@aperainst.de
www.aperainst.de

Compatible models for EC60-DE replacement probe:

● **EC60-Z** ●**PC60-Z**

How to replace the probe?

Screw off the probe ring, unplug the probe, plug in the new replacement probe (pay attention to the probe's position), and screw on the probe ring.



**Video tutorials are available at www.aperainst.de or on YouTube (Apera Instruments Europe Official Channel)

We warrant this electrode to be free from defects in material and workmanship and agree to repair or replace free of charge, at option of APERA INSTRUMENTS (Europe) GmbH any malfunctioned or damaged product attributable to responsibility of APERA INSTRUMENTS (Europe) GmbH for a period of six-month. This warranty does not apply to defects resulting from actions such as misuse (violation of the instructions in this manual or operations in the manner not specified in this manual), improper maintenance, or unauthorized repairs. Warranty period is the time limit to provide free service for the products purchased by customers, not the service life of the probe.

APERA INSTRUMENTS (Europe) GmbH
Wilhelm-Muthmann-Straße 15, 42329 Wuppertal, Germany
Tel. +49 202 51988998
info@aperainst.de
www.aperainst.de